



Universidad de Valladolid



CERTIFICATE OF PRODUCT CYTOTOXICITY

The University of Valladolid, Valladolid, Spain, through its Eye Institute (Instituto Universitario de Oftalmobiología Aplicada IOBA) which has implemented and maintains a Quality Management System which fulfills the requirements of the ISO-9001:2008 standards (valid up to 17-04-2018) registration number: ES-0101/2006.

CERTIFIES THAT:

lot(s)/batch(es) number(s): 141125; 551477; 558231; 590929; 591057;

The product: PFCL Medical Devices

930766; 950418; 952216

provided by: ARCADOPHTA, 11 Rue Antoine Ricord, 31100, Toulouse, France. VAT: FR 54440530558.

"does not produce cytotoxicity"

on cultures of **retinal pigment epithelium** (RPE) cells from **human** origin (ARPE-19 cell line), and **neuroretina** explants from **porcine** origin in our experimental conditions performed following the UNE ISO guidelines (UNE-EN ISO 10993-5: 2009, 10993-12: 2012).

Signed:

Sponsor

Signature:

Mr Gerard Reboul technical director

Date: October 24, 2016

Innovation and Transference Vicerector

Signature:

Prof. Dr. Celedonio Álvarez-González

Date: October 24, 2016

Head of Test Facility Management
(Director of IOBA)

Signature:

Prof. Dr. Miguel J. Maldonado-López

Date: October 24, 2016

Study Director

Signature:

Dr. Girish Kumar-Srivatava

Date: October 24, 2016

Quality Assurance Manager

Signature:

D. Miguel Ángel García-Hontoria

Date: October 24, 2016

Head of Retina Group

Signature:

Prof. Dr. J. Carlos Pastor-Jimeno

Date: October 24, 2016

